

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050720\
 Data File : VN061339.D
 Acq On : 7 May 2020 17:16
 Operator : JC/MD
 Sample : L2548-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ST008MW186-200504

Quant Time: May 08 03:36:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 05 07:01:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	91332	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	165799	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	160339	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	65807	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	76334	52.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.64%	
35) Dibromofluoromethane	7.55	113	51722	52.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.52%	
50) Toluene-d8	10.06	98	207194	51.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.68%	
62) 4-Bromofluorobenzene	12.38	95	73029	48.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.12%	
Target Compounds						
21) trans-1,2-Dichloroethene	4.99	96	1341	1.278	ug/l #	79
27) cis-1,2-Dichloroethene	6.78	96	12444	10.106	ug/l	90
44) Trichloroethene	8.80	130	4151	3.105	ug/l	90
64) Tetrachloroethene	10.60	164	32310	23.266	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050720\
Data File : VN061339.D
Acq On : 7 May 2020 17:16
Operator : JC/MD
Sample : L2548-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ST008MW186-200504

Quant Time: May 08 03:36:16 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Tue May 05 07:01:08 2020
Response via : Initial Calibration

