

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050720\
 Data File : VN061341.D
 Acq On : 7 May 2020 18:02
 Operator : JC/MD
 Sample : L2548-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SS036MW004-200504

Quant Time: May 08 03:42:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 05 07:01:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	92055	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	168144	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	160392	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	68631	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	78803	54.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.20%	
35) Dibromofluoromethane	7.55	113	52190	52.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.98%	
50) Toluene-d8	10.06	98	211369	52.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.30%	
62) 4-Bromofluorobenzene	12.38	95	75534	49.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.04%	
Target Compounds						
27) cis-1,2-Dichloroethene	6.78	96	2041	1.644	ug/l	87
44) Trichloroethene	8.80	130	3382	2.495	ug/l	98
64) Tetrachloroethene	10.60	164	14257	10.263	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN050720\
Data File : VN061341.D
Acq On : 7 May 2020 18:02
Operator : JC/MD
Sample : L2548-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SS036MW004-200504

Quant Time: May 08 03:42:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Tue May 05 07:01:08 2020
Response via : Initial Calibration

