

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050720\
 Data File : VN061342.D
 Acq On : 7 May 2020 18:25
 Operator : JC/MD
 Sample : L2548-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Z3-DUP-0549-200504

Quant Time: May 08 03:45:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 05 07:01:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	88612	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	160988	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	157751	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	66949	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	77185	55.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.10%	
35) Dibromofluoromethane	7.55	113	50587	53.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.28%	
50) Toluene-d8	10.06	98	203353	52.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.80%	
62) 4-Bromofluorobenzene	12.38	95	72852	49.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.78%	
Target Compounds						
27) cis-1,2-Dichloroethene	6.78	96	2026	1.696	ug/l	85
44) Trichloroethene	8.80	130	3443	2.653	ug/l	99
64) Tetrachloroethene	10.60	164	14647	10.720	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050720\
Data File : VN061342.D
Acq On : 7 May 2020 18:25
Operator : JC/MD
Sample : L2548-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Z3-DUP-0549-200504

Quant Time: May 08 03:45:47 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Tue May 05 07:01:08 2020
Response via : Initial Calibration

