

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050721\
 Data File : VN067032.D
 Acq On : 7 May 2021 1:11
 Operator : JC/MD
 Sample : M2307-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-18

Manual Integrations
 APPROVED

MMDadoda
 5/7/2021 6:23:53 PM

Quant Time: May 07 07:32:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N050721W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri May 07 07:26:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.115	128	78615	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.518	114	393120	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.352	117	315742	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.954	65	203385	29.52	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	98.40%
60) 4-Bromofluorobenzene	12.355	95	133989m	25.51	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	85.03%
63) Toluene-d8	10.030	98	503181	32.17	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	107.23%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050721\
Data File : VN067032.D
Acq On : 7 May 2021 1:11
Operator : JC/MD
Sample : M2307-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-18

Manual Integrations
APPROVED

MMDadoda
5/7/2021 6:23:53 PM

Quant Time: May 07 07:32:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N050721W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri May 07 07:26:20 2021
Response via : Initial Calibration

