

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050721\
 Data File : VN067034.D
 Acq On : 7 May 2021 1:59
 Operator : JC/MD
 Sample : M2257-02 5X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 14B-2

Manual Integrations
 APPROVED

MMDadoda
 5/7/2021 6:23:57 PM

Quant Time: May 07 07:32:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N050721W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri May 07 07:26:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.115	128	78665	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.520	114	385302	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.352	117	312125	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.954	65	207370	30.08	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery = 100.27%			
60) 4-Bromofluorobenzene	12.358	95	136517m	26.29	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery = 87.63%			
63) Toluene-d8	10.030	98	496394	32.11	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery = 107.03%			
Target Compounds						Qvalue
18) Methylene Chloride	4.473	84	6480m	1.07	ug/l	
25) Chloroform	7.300	83	9215m	0.89	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050721\
Data File : VN067034.D
Acq On : 7 May 2021 1:59
Operator : JC/MD
Sample : M2257-02 5X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
14B-2

Manual Integrations
APPROVED

MMDadoda
5/7/2021 6:23:57 PM

Quant Time: May 07 07:32:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N050721W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri May 07 07:26:20 2021
Response via : Initial Calibration

