

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050721\
 Data File : VN067040.D
 Acq On : 7 May 2021 4:22
 Operator : JC/MD
 Sample : M2274-04 5X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 14B-4

Manual Integrations
 APPROVED

MMDadoda
 5/7/2021 6:24:28 PM

Quant Time: May 07 07:33:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N050721W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri May 07 07:26:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.120	128	71806	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.520	114	331593	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.352	117	267902	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.954	65	188764	30.00	ug/l	0.00
Spiked Amount 30.000	Range	91 - 110	Recovery	=	100.00%	
60) 4-Bromofluorobenzene	12.358	95	123102m	27.62	ug/l	0.00
Spiked Amount 30.000	Range	63 - 112	Recovery	=	92.07%	
63) Toluene-d8	10.030	98	429479	32.36	ug/l	0.00
Spiked Amount 30.000	Range	91 - 112	Recovery	=	107.87%	
Target Compounds						
18) Methylene Chloride	4.481	84	5153m	0.93	ug/l	Qvalue
25) Chloroform	7.297	83	9878	1.05	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050721\
Data File : VN067040.D
Acq On : 7 May 2021 4:22
Operator : JC/MD
Sample : M2274-04 5X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
14B-4

Manual Integrations
APPROVED

MMDadoda
5/7/2021 6:24:28 PM

Quant Time: May 07 07:33:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N050721W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri May 07 07:26:20 2021
Response via : Initial Calibration

