


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050721\
 Data File : VN067018.D
 Acq On : 6 May 2021 18:24
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N050721W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Fri May 07 07:26:20 2021



AutoFind: Scans 3952, 3953, 3954; Background Corrected with Scan 3936

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.1	5612	PASS
75	95	30	60	52.1	15285	PASS
95	95	100	100	100.0	29325	PASS
96	95	5	9	6.4	1868	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	73.3	21507	PASS
175	174	5	9	7.8	1685	PASS
176	174	95	101	94.5	20333	FAIL*
177	176	5	9	7.4	1502	PASS