

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050724\
 Data File : VN082018.D
 Acq On : 07 May 2024 11:14
 Operator : JC\MD
 Sample : VN0507WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0507WBL01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/08/2024
 Supervised By : Mahesh Dadoda 05/08/2024

Quant Time: May 08 07:58:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Apr 27 04:37:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.818	128	30328m	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	182222	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	155279	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.577	65	80379	31.655	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	105.533%
60) 4-Bromofluorobenzene	12.847	95	70187	27.160	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	90.533%
63) Toluene-d8	10.571	98	229989	31.188	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	103.967%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050724\
Data File : VN082018.D
Acq On : 07 May 2024 11:14
Operator : JC\MD
Sample : VN0507WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0507WBL01

Quant Time: May 08 07:58:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Apr 27 04:37:27 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/08/2024
Supervised By :Mahesh Dadoda 05/08/2024

