



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050724\
Data File : VN082014.D
Acq On : 07 May 2024 08:25
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Sat Apr 27 04:37:27 2024



AutoFind: Scans 1852, 1853, 1854; Background Corrected with Scan 1845

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.0	8142	PASS
75	95	30	60	55.7	20562	PASS
95	95	100	100	100.0	36947	PASS
96	95	5	9	6.9	2549	PASS
173	174	0.00	2	0.3	62	PASS
174	95	50	100	60.7	22437	PASS
175	174	5	9	9.1	2045	FAIL*
176	174	95	101	95.6	21445	PASS
177	176	5	9	6.9	1487	PASS