

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050820\
 Data File : VN061360.D
 Acq On : 8 May 2020 14:36
 Operator : JC/MD
 Sample : L2544-11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SS038MW009-200505

Quant Time: May 11 02:54:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 05 07:01:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	99710	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	181963	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	176327	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	82501	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	82658	52.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.78%	
35) Dibromofluoromethane	7.55	113	53822	50.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.04%	
50) Toluene-d8	10.06	98	227007	51.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.50%	
62) 4-Bromofluorobenzene	12.38	95	91311	55.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.64%	
Target Compounds						
16) Acetone	3.78	43	8688	8.911	ug/l	Qvalue 100
39) Methylcyclohexane	9.04	83	1058	0.518	ug/l #	91
40) Benzene	8.00	78	19916	3.670	ug/l	98
65) Chlorobenzene	11.41	112	1009906	288.730	ug/l	99
73) Isopropylbenzene	12.22	105	13515	2.228	ug/l	98
78) n-propylbenzene	12.56	91	8400	1.168	ug/l	94
83) tert-Butylbenzene	12.97	119	22296	5.268	ug/l	92
85) sec-Butylbenzene	13.15	105	26783	4.641	ug/l #	65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050820\
Data File : VN061360.D
Acq On : 8 May 2020 14:36
Operator : JC/MD
Sample : L2544-11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SS038MW009-200505

Quant Time: May 11 02:54:52 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Tue May 05 07:01:08 2020
Response via : Initial Calibration

