

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050820\
 Data File : VN061363.D
 Acq On : 8 May 2020 15:46
 Operator : JC/MD
 Sample : L2544-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Z3-DUP-0551-200505

Quant Time: May 11 04:12:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 05 07:01:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	103639	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	188194	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	181588	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	84796	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	84678	51.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.28%	
35) Dibromofluoromethane	7.55	113	53510	48.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.16%	
50) Toluene-d8	10.06	98	241052	53.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.28%	
62) 4-Bromofluorobenzene	12.38	95	94179	55.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.34%	
Target Compounds						
65) Chlorobenzene	11.41	112	65256	18.116	ug/l	99
73) Isopropylbenzene	12.22	105	26353	4.227	ug/l	100
83) tert-Butylbenzene	12.97	119	33287	7.651	ug/l	93
85) sec-Butylbenzene	13.15	105	72279	12.186	ug/l #	36
87) 1,3-Dichlorobenzene	13.26	146	1336	0.504	ug/l #	66
88) 1,4-Dichlorobenzene	13.34	146	2959	1.093	ug/l #	59

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050820\
Data File : VN061363.D
Acq On : 8 May 2020 15:46
Operator : JC/MD
Sample : L2544-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Z3-DUP-0551-200505

Quant Time: May 11 04:12:35 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Tue May 05 07:01:08 2020
Response via : Initial Calibration

