

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050820\
 Data File : VN061367.D
 Acq On : 8 May 2020 17:19
 Operator : JC/MD
 Sample : PB128729ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB128729ZHE#03

Quant Time: May 11 04:26:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 05 07:01:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	96917	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	179018	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	178826	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	73568	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	81867	53.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.76%	
35) Dibromofluoromethane	7.55	113	49877	47.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.24%	
50) Toluene-d8	10.06	98	230513	53.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.84%	
62) 4-Bromofluorobenzene	12.38	95	84783	52.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.42%	
Target Compounds						
16) Acetone	3.78	43	9619	10.935	ug/l	97
20) Methylene Chloride	4.51	84	1550	1.296	ug/l	93
43) Isopropyl Acetate	8.13	43	38012	10.902	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN050820\
Data File : VN061367.D
Acq On : 8 May 2020 17:19
Operator : JC/MD
Sample : PB128729ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB128729ZHE#03

Quant Time: May 11 04:26:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Tue May 05 07:01:08 2020
Response via : Initial Calibration

