

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050820\
 Data File : VN061377.D
 Acq On : 8 May 2020 21:11
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: May 11 04:50:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 05 07:01:08 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	0.530	0.468	11.7	73	0.00
3 P	Chloromethane	0.642	0.637	0.8	86	0.00
4 C	Vinyl Chloride	0.729	0.709	2.7#	83	0.00
5 T	Bromomethane	0.361	0.281	22.2	68	0.00
6 T	Chloroethane	0.462	0.426	7.8	80	0.00
7 T	Trichlorofluoromethane	0.823	0.892	-8.4	96	0.00
8 T	Diethyl Ether	0.343	0.357	-4.1	88	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.530	0.522	1.5	87	0.00
10 T	Methyl Iodide	0.681	0.503	26.1#	65	0.00
11 T	Tert butyl alcohol	0.106	0.117	-10.4	100	0.00
12 CM	1,1-Dichloroethene	0.528	0.493	6.6#	83	0.00
13 T	Acrolein	0.071	0.059	16.9	77	0.00
14 T	Allyl chloride	0.978	1.081	-10.5	91	0.00
15 T	Acrylonitrile	0.325	0.351	-8.0	93	0.00
16 T	Acetone	0.398	0.316	20.6	94	0.00
17 T	Carbon Disulfide	1.457	0.955	34.5#	56	0.00
18 T	Methyl Acetate	1.087	0.986	9.3	93	0.00
19 T	Methyl tert-butyl Ether	2.016	1.983	1.6	88	0.00
20 T	Methylene Chloride	0.617	0.588	4.7	85	0.00
21 T	trans-1,2-Dichloroethene	0.574	0.538	6.3	81	0.00
22 T	Diisopropyl ether	2.268	2.343	-3.3	91	0.00
23 T	Vinyl Acetate	1.826	1.872	-2.5	88	0.00
24 P	1,1-Dichloroethane	1.185	1.180	0.4	88	0.00
25 T	2-Butanone	0.468	0.500	-6.8	92	0.00
26 T	2,2-Dichloropropane	0.998	0.912	8.6	86	0.00
27 T	cis-1,2-Dichloroethene	0.674	0.650	3.6	85	0.00
28 T	Bromochloromethane	0.562	0.582	-3.6	98	0.00
29 T	Tetrahydrofuran	0.315	0.337	-7.0	92	0.00
30 C	Chloroform	1.201	1.163	3.2#	86	0.00
31 T	Cyclohexane	1.151	0.964	16.2	77	0.00
32 T	1,1,1-Trichloroethane	1.046	0.998	4.6	84	0.00
33 S	1,2-Dichloroethane-d4	0.791	0.810	-2.4	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	86	0.00
35 S	Dibromofluoromethane	0.296	0.263	11.1	79	0.00
36 T	1,1-Dichloropropene	0.501	0.481	4.0	82	0.00
37 T	Ethyl Acetate	0.587	0.606	-3.2	89	0.00
38 T	Carbon Tetrachloride	0.400	0.444	-11.0	89	0.00
39 T	Methylcyclohexane	0.561	0.478	14.8	72	0.00
40 TM	Benzene	1.491	1.423	4.6	83	0.00
41 T	Methacrylonitrile	0.244	0.290	-18.9	107	0.00
42 TM	1,2-Dichloroethane	0.565	0.570	-0.9	87	0.00
43 T	Isopropyl Acetate	0.974	0.993	-2.0	88	0.00
44 TM	Trichloroethene	0.403	0.351	12.9	77	0.00
45 C	1,2-Dichloropropane	0.398	0.404	-1.5#	88	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050820\
 Data File : VN061377.D
 Acq On : 8 May 2020 21:11
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: May 11 04:50:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 05 07:01:08 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.247	0.249	-0.8	85	0.00
47 T	Bromodichloromethane	0.529	0.541	-2.3	87	0.00
48 T	Methyl methacrylate	0.439	0.472	-7.5	90	0.00
49 T	1,4-Dioxane	0.005	0.005	0.0	102	0.00
50 S	Toluene-d8	1.205	1.226	-1.7	91	0.00
51 T	4-Methyl-2-Pentanone	0.573	0.621	-8.4	90	0.00
52 CM	Toluene	0.898	0.873	2.8#	82	0.00
53 T	t-1,3-Dichloropropene	0.597	0.588	1.5	84	0.00
54 T	cis-1,3-Dichloropropene	0.631	0.616	2.4	85	0.00
55 T	1,1,2-Trichloroethane	0.358	0.355	0.8	86	0.00
56 T	Ethyl methacrylate	0.540	0.585	-8.3	86	0.00
57 T	1,3-Dichloropropane	0.630	0.635	-0.8	85	0.00
58 T	2-Chloroethyl Vinyl ether	0.290	0.293	-1.0	86	0.00
59 T	2-Hexanone	0.417	0.456	-9.4	90	0.00
60 T	Dibromochloromethane	0.375	0.380	-1.3	86	0.00
61 T	1,2-Dibromoethane	0.359	0.353	1.7	82	0.00
62 S	4-Bromofluorobenzene	0.454	0.466	-2.6	92	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	85	0.00
64 T	Tetrachloroethene	0.433	0.362	16.4	72	0.00
65 PM	Chlorobenzene	0.992	0.969	2.3	83	0.00
66 T	1,1,1,2-Tetrachloroethane	0.355	0.363	-2.3	85	0.00
67 C	Ethyl Benzene	1.857	1.824	1.8#	81	0.00
68 T	m/p-Xylenes	0.680	0.665	2.2	80	0.00
69 T	o-Xylene	0.653	0.638	2.3	81	0.00
70 T	Styrene	1.093	1.112	-1.7	82	0.00
71 P	Bromoform	0.261	0.282	-8.0	87	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
73 T	Isopropylbenzene	3.676	3.541	3.7	81	0.00
74 T	N-amyl acetate	1.861	1.885	-1.3	84	0.00
75 P	1,1,2,2-Tetrachloroethane	1.146	1.180	-3.0	88	0.00
76 T	1,2,3-Trichloropropane	1.054	1.075	-2.0	88	0.00
77 T	Bromobenzene	0.849	0.827	2.6	82	0.00
78 T	n-propylbenzene	4.360	4.232	2.9	80	0.00
79 T	2-Chlorotoluene	2.563	2.494	2.7	82	0.00
80 T	1,3,5-Trimethylbenzene	3.042	2.976	2.2	81	0.00
81 T	trans-1,4-Dichloro-2-butene	0.415	0.414	0.2	85	0.00
82 T	4-Chlorotoluene	2.701	2.671	1.1	83	0.00
83 T	tert-Butylbenzene	2.565	2.431	5.2	78	0.00
84 T	1,2,4-Trimethylbenzene	3.134	3.043	2.9	81	0.00
85 T	sec-Butylbenzene	3.497	3.279	6.2	77	0.00
86 T	p-Isopropyltoluene	3.151	2.929	7.0	76	0.00
87 T	1,3-Dichlorobenzene	1.562	1.512	3.2	81	0.00
88 T	1,4-Dichlorobenzene	1.596	1.512	5.3	81	0.00
89 T	n-Butylbenzene	2.889	2.668	7.6	75	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050820\
Data File : VN061377.D
Acq On : 8 May 2020 21:11
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: May 11 04:50:46 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Tue May 05 07:01:08 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.281	0.395	-40.6#	115	0.00
91 T	1,2-Dichlorobenzene	1.498	1.451	3.1	81	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.242	0.249	-2.9	83	0.00
93 T	1,2,4-Trichlorobenzene	0.912	0.830	9.0	75	0.00
94 T	Hexachlorobutadiene	0.381	0.356	6.6	77	0.00
95 T	Naphthalene	2.970	2.721	8.4	74	0.00
96 T	1,2,3-Trichlorobenzene	0.900	0.824	8.4	76	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6