

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050821\
 Data File : VN067049.D
 Acq On : 7 May 2021 14:19
 Operator : JC/MD
 Sample : VN0508WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0508WBL01

Quant Time: May 07 18:47:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N050721W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri May 07 07:26:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.112	128	99430	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.517	114	521841	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.350	117	453109	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.952	65	262063	30.08	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	100.27%
60) 4-Bromofluorobenzene	12.353	95	156178	20.72	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	69.07%
63) Toluene-d8	10.027	98	680603	30.32	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.07%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050821\
Data File : VN067049.D
Acq On : 7 May 2021 14:19
Operator : JC/MD
Sample : VN0508WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0508WBL01

Quant Time: May 07 18:47:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N050721W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri May 07 07:26:20 2021
Response via : Initial Calibration

