

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050821\
 Data File : VN067054.D
 Acq On : 7 May 2021 16:28
 Operator : JC/MD
 Sample : M2296-01
 Misc : 7.30g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP-1

Quant Time: May 07 16:59:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.581	168	420211	50.00	ug/l	#-0.01
34) 1,4-Difluorobenzene	8.515	114	708427	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.352	117	594706	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.297	152	248769	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.946	65	352456	53.29	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 106.58%	
35) Dibromofluoromethane	7.506	113	246356	51.21	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 102.42%	
50) Toluene-d8	10.027	98	950325	51.87	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 103.74%	
62) 4-Bromofluorobenzene	12.355	95	254547	41.03	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 82.06%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050821\
Data File : VN067054.D
Acq On : 7 May 2021 16:28
Operator : JC/MD
Sample : M2296-01
Misc : 7.30g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP-1

Quant Time: May 07 16:59:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

