

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050821\
 Data File : VN067058.D
 Acq On : 7 May 2021 18:10
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: May 08 03:27:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.592	168	419228	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.520	114	703942	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.352	117	587789	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.294	152	209509	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.954	65	360314	54.60	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 109.20%	
35) Dibromofluoromethane	7.514	113	255669	53.49	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 106.98%	
50) Toluene-d8	10.030	98	959813	52.73	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.46%	
62) 4-Bromofluorobenzene	12.353	95	270740	43.92	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 87.84%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050821\
Data File : VN067058.D
Acq On : 7 May 2021 18:10
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: May 08 03:27:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

