

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050824\
 Data File : VN082033.D
 Acq On : 08 May 2024 17:37
 Operator : JC\MD
 Sample : P2424-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 240507002-04-TRIP-BLANK

Quant Time: May 09 01:04:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Apr 27 04:37:27 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/09/2024
 Supervised By : Mahesh Dadoda 05/09/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.812	128	29066	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	169940	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	141823	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.583	65	75441m	31.000	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	103.333%
60) 4-Bromofluorobenzene	12.853	95	66141	28.023	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	93.400%
63) Toluene-d8	10.571	98	209100	31.046	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	103.500%
Target Compounds					Qvalue	
25) Chloroform	7.965	83	17658	5.030	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050824\
Data File : VN082033.D
Acq On : 08 May 2024 17:37
Operator : JC\MD
Sample : P2424-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
240507002-04-TRIP-BLANK

Quant Time: May 09 01:04:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Apr 27 04:37:27 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/09/2024
Supervised By :Mahesh Dadoda 05/09/2024

