

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050824\
 Data File : VN082037.D
 Acq On : 08 May 2024 19:13
 Operator : JC\MD
 Sample : P2427-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 240424030-03-TRIP-BLANK

Quant Time: May 09 01:06:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Apr 27 04:37:27 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/09/2024
 Supervised By : Mahesh Dadoda 05/09/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.818	128	25600	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.106	114	151338	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	130241	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.583	65	69307	32.336	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	107.800%
60) 4-Bromofluorobenzene	12.847	95	56638	26.131	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	87.100%
63) Toluene-d8	10.571	98	189897	30.702	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	102.333%
Target Compounds						
25) Chloroform	7.965	83	4324	1.399	ug/l	84
55) 2-Chloroethyl vinyl ether	10.153	63	1972m	1.300	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050824\
Data File : VN082037.D
Acq On : 08 May 2024 19:13
Operator : JC\MD
Sample : P2427-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
240424030-03-TRIP-BLANK

Quant Time: May 09 01:06:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Apr 27 04:37:27 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 05/09/2024
Supervised By : Mahesh Dadoda 05/09/2024

