

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050825\
 Data File : VN086544.D
 Acq On : 08 May 2025 14:36
 Operator : JC\MD
 Sample : PB167895ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167895ZHE#02

Quant Time: May 09 01:52:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	202516	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	369362	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	351871	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	153642	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	134814	45.907	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	91.820%		
35) Dibromofluoromethane	8.165	113	105631	61.619	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	123.240%		
50) Toluene-d8	10.565	98	466275	50.893	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	101.780%		
62) 4-Bromofluorobenzene	12.847	95	168476	50.416	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	100.840%		
Target Compounds						
16) Acetone	4.430	43	46813	41.343	ug/l	98
20) Methylene Chloride	5.277	84	7925	2.919	ug/l	91
43) Isopropyl Acetate	8.688	43	29755	2.788	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050825\
Data File : VN086544.D
Acq On : 08 May 2025 14:36
Operator : JC\MD
Sample : PB167895ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167895ZHE#02

Quant Time: May 09 01:52:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

