

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050825\
 Data File : VN086547.D
 Acq On : 08 May 2025 15:49
 Operator : JC\MD
 Sample : PB167895ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167895ZHE#05

Quant Time: May 09 01:53:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	216290	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	393667	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	383817	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	167550	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	142002	45.276	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.560%
35) Dibromofluoromethane	8.165	113	112872	61.778	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	123.560%
50) Toluene-d8	10.565	98	505188	51.736	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.480%
62) 4-Bromofluorobenzene	12.847	95	184129	51.698	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.400%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	33010	25.312	ug/l	92
20) Methylene Chloride	5.271	84	5163	1.781	ug/l #	76
43) Isopropyl Acetate	8.683	43	34841	3.249	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050825\
Data File : VN086547.D
Acq On : 08 May 2025 15:49
Operator : JC\MD
Sample : PB167895ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167895ZHE#05

Quant Time: May 09 01:53:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

