

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN050919\
 Data File : VN055486.D
 Acq On : 9 May 2019 10:34
 Operator : JC/SP
 Sample : VN0509MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0509MBL01

Quant Time: May 09 15:33:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N050819W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 04:17:44 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	833103	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1406360	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1158339	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	302866	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	579354	46.19	ug/l	0.00
Spiked Amount			Recovery	=	92.38%	
35) Dibromofluoromethane	7.59	113	447798	49.36	ug/l	0.00
Spiked Amount			Recovery	=	98.72%	
50) Toluene-d8	10.09	98	1687938	47.74	ug/l	0.00
Spiked Amount			Recovery	=	95.48%	
62) 4-Bromofluorobenzene	12.40	95	474694	39.82	ug/l	0.00
Spiked Amount			Recovery	=	79.64%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN050919\
Data File : VN055486.D
Acq On : 9 May 2019 10:34
Operator : JC/SP
Sample : VN0509MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0509MBL01

Quant Time: May 09 15:33:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N050819W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 04:17:44 2019
Response via : Initial Calibration

