

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050919\
 Data File : VN055518.D
 Acq On : 10 May 2019 00:49
 Operator : JC/SP
 Sample : K2758-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CKMW-23

Quant Time: May 10 13:50:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N050819W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 04:17:44 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	899870	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1494993	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1259551	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	344992	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	610718	45.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.14%	
35) Dibromofluoromethane	7.59	113	476013	49.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.72%	
50) Toluene-d8	10.09	98	1846957	49.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.30%	
62) 4-Bromofluorobenzene	12.40	95	528883	41.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.48%	
Target Compounds						
19) Methyl tert-butyl Ether	5.05	73	27489	0.841	ug/l	94
27) cis-1,2-Dichloroethene	6.83	96	105659	8.574	ug/l	92
30) Chloroform	7.38	83	6746	0.330	ug/l	88
44) Trichloroethene	8.83	130	74464	6.846	ug/l	93
64) Tetrachloroethene	10.63	164	279884	27.118	ug/l	99
89) n-Butylbenzene	13.62	91	6549	0.330	ug/l	86
95) Naphthalene	15.14	128	13162	3.798	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN050919\
Data File : VN055518.D
Acq On : 10 May 2019 00:49
Operator : JC/SP
Sample : K2758-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CKMW-23

Quant Time: May 10 13:50:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N050819W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 04:17:44 2019
Response via : Initial Calibration

