

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050919\
 Data File : VN055526.D
 Acq On : 10 May 2019 4:09
 Operator : JC/SP
 Sample : K2758-23
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CKMW-22

Quant Time: May 10 13:54:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N050819W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 04:17:44 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	851030	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1429724	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1196392	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	326589	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	591089	46.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.26%	
35) Dibromofluoromethane	7.59	113	452199	49.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.06%	
50) Toluene-d8	10.09	98	1757787	48.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.82%	
62) 4-Bromofluorobenzene	12.40	95	498142	41.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.22%	
Target Compounds						
30) Chloroform	7.37	83	24301	1.258	ug/l	97
64) Tetrachloroethene	10.63	164	5051	0.515	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN050919\
Data File : VN055526.D
Acq On : 10 May 2019 4:09
Operator : JC/SP
Sample : K2758-23
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CKMW-22

Quant Time: May 10 13:54:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N050819W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 04:17:44 2019
Response via : Initial Calibration

