

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN050919\
 Data File : VN055495.D
 Acq On : 9 May 2019 14:23
 Operator : JC/SP
 Sample : K2673-01
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-11-1.5-2.0

Quant Time: May 09 15:37:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N050819W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 04:17:44 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	798073	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1312694	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1134272	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	423944	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	549850	45.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.52%	
35) Dibromofluoromethane	7.59	113	408368	48.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.46%	
50) Toluene-d8	10.09	98	1622725	49.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.36%	
62) 4-Bromofluorobenzene	12.40	95	533090	47.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.82%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN050919\
Data File : VN055495.D
Acq On : 9 May 2019 14:23
Operator : JC/SP
Sample : K2673-01
Misc : 5.01µL/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-11-1.5-2.0

Quant Time: May 09 15:37:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N050819W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 04:17:44 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.66 min Scan# 1832

Delta R.T. -0.00 min

Lab File: VN055495.D

Acq: 9 May 2019 14:23

Instrument :

MSVOA_N

ClientSampleId :


SB-11-1.5-2.0

Tot Ion:168 Resp: 798073

Ion Ratio Lower Upper




168 100

99 57.3 47.0 70.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO



#33

1,2-Dichloroethane-d4

Concen: 45.759 ug/l

RT: 8.03 min Scan# 1945

Delta R.T. -0.00 min

Lab File: VN055495.D


Acq: 9 May 2019 14:23

Tgt Ion: 65 Resp: 549850

Ion Ratio Lower Upper



65 100


67 54.0 0.0 109.6



Abundance Ion 64.90 (64.60 to 65.60): VNO

Ion 66.90 (66.60 to 67.60): VNO









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.58 min Scan# 2118
 Delta R.T. -0.00 min
 Lab File: VN055495.D
 Acq: 9 May 2019 14:23

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-11-1.5-2.0

Tot Ion:114 Resp: 1312694
 Ion Ratio Lower Upper
 114 100
 63 23.7 0.0 47.4
 88 16.2 0.0 34.0




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): V
 Ion 87.90 (87.60 to 88.60): V






#35
 Dibromofluoromethane
 Concen: 48.225 ug/l
 RT: 7.59 min Scan# 1809
 Delta R.T. -0.00 min
 Lab File: VN055495.D
 Acq: 9 May 2019 14:23

Tgt Ion:113 Resp: 408368
 Ion Ratio Lower Upper
 113 100
 111 101.7 82.9 124.3
 192 19.3 15.6 23.4



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 49.175 ug/l

RT: 10.09 min Scan# 2587

Delta R.T. -0.00 min

Lab File: VN055495.D

Acq: 9 May 2019 14:23

Instrument :

MSVOA_N

ClientSampleId :





SB-11-1.5-2.0

Tot Ion: 98 Resp: 1622725

Ion Ratio Lower Upper

98 100

100 63.8 50.4 75.6



#62

4-Bromofluorobenzene

Concen: 47.911 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.00 min

Lab File: VN055495.D

Acq: 9 May 2019 14:23




Tgt Ion: 95 Resp: 533090


Ion Ratio Lower Upper

95 100

174 77.6 0.0 152.2

176 74.8 0.0 147.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. -0.00 min
Lab File: VN055495.D
Acq: 9 May 2019 14:23

Instrument :
MSVOA_N
ClientSampleId :
SB-11-1.5-2.0

Tot Ion:117 Resp: 1134272
Ion Ratio Lower Upper
117 100
82 59.1 48.0 72.0
119 31.9 26.0 39.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.34 min Scan# 3598
Delta R.T. -0.00 min
Lab File: VN055495.D
Acq: 9 May 2019 14:23

Tgt Ion:152 Resp: 423944
Ion Ratio Lower Upper
152 100
115 60.3 30.1 90.5
150 157.2 0.0 355.0

