

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050922\
 Data File : VN072377.D
 Acq On : 09 May 2022 14:21
 Operator : JC\MD
 Sample : N2740-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: May 10 06:45:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	289032	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	546812	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	585839	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	228722	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	202477	45.631	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.260%
35) Dibromofluoromethane	8.016	113	164741	47.070	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.140%
50) Toluene-d8	10.433	98	693294	50.412	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.820%
62) 4-Bromofluorobenzene	12.727	95	271069	59.383	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	118.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050922\
Data File : VN072377.D
Acq On : 09 May 2022 14:21
Operator : JC\MD
Sample : N2740-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: May 10 06:45:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

