

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050922\
 Data File : VN072372.D
 Acq On : 09 May 2022 11:41
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: May 10 06:42:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	88	0.00
2 T	Dichlorodifluoromethane	0.621	0.485	21.9	81	0.00
3 P	Chloromethane	0.762	0.533	30.1#	78	0.00
4 C	Vinyl Chloride	0.563	0.480	14.7#	85	0.00
5 T	Bromomethane	0.307	0.270	12.1	91	-0.01
6 T	Chloroethane	0.372	0.296	20.4	76	0.00
7 T	Trichlorofluoromethane	0.983	0.864	12.1	87	0.00
8 T	Diethyl Ether	0.356	0.369	-3.7	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.517	0.516	0.2	97	0.00
10 T	Methyl Iodide	0.627	0.639	-1.9	93	0.00
11 T	Tert butyl alcohol	0.126	0.132	-4.8	96	0.00
12 CM	1,1-Dichloroethene	0.478	0.458	4.2#	90	0.00
13 T	Acrolein	0.127	0.119	6.3	89	0.00
14 T	Allyl chloride	0.992	1.005	-1.3	94	0.00
15 T	Acrylonitrile	0.381	0.381	0.0	90	0.00
16 T	Acetone	0.331	0.390	-17.8	113	0.00
17 T	Carbon Disulfide	1.222	0.906	25.9#	76	0.00
18 T	Methyl Acetate	1.039	0.972	6.4	97	0.00
19 T	Methyl tert-butyl Ether	1.872	2.031	-8.5	97	0.00
20 T	Methylene Chloride	0.634	0.591	6.8	93	0.01
21 T	trans-1,2-Dichloroethene	0.527	0.484	8.2	89	0.00
22 T	Diisopropyl ether	2.042	2.139	-4.8	98	0.00
23 T	Vinyl Acetate	1.715	1.811	-5.6	95	0.00
24 P	1,1-Dichloroethane	1.142	1.093	4.3	91	0.00
25 T	2-Butanone	0.527	0.557	-5.7	98	0.00
26 T	2,2-Dichloropropane	0.892	0.963	-8.0	101	0.00
27 T	cis-1,2-Dichloroethene	0.642	0.643	-0.2	96	0.00
28 T	Bromochloromethane	0.512	0.497	2.9	88	0.00
29 T	Tetrahydrofuran	0.346	0.346	0.0	91	0.00
30 C	Chloroform	1.147	1.105	3.7#	95	0.00
31 T	Cyclohexane	1.026	0.885	13.7	86	0.00
32 T	1,1,1-Trichloroethane	0.976	0.983	-0.7	95	0.00
33 S	1,2-Dichloroethane-d4	0.768	0.775	-0.9	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	86	0.00
35 S	Dibromofluoromethane	0.320	0.358	-11.9	99	0.00
36 T	1,1-Dichloropropene	0.466	0.466	0.0	90	0.00
37 T	Ethyl Acetate	0.619	0.632	-2.1	92	0.00
38 T	Carbon Tetrachloride	0.499	0.504	-1.0	95	0.00
39 T	Methylcyclohexane	0.521	0.539	-3.5	90	0.00
40 TM	Benzene	1.445	1.431	1.0	91	0.00
41 T	Methacrylonitrile	0.313	0.323	-3.2	89	0.00
42 TM	1,2-Dichloroethane	0.553	0.558	-0.9	93	0.00
43 T	Isopropyl Acetate	1.015	1.012	0.3	96	0.00
44 TM	Trichloroethene	0.336	0.354	-5.4	98	0.00
45 C	1,2-Dichloropropane	0.390	0.400	-2.6#	94	0.00
46 T	Dibromomethane	0.245	0.256	-4.5	95	0.00
47 T	Bromodichloromethane	0.513	0.554	-8.0	96	0.00
48 T	Methyl methacrylate	0.409	0.454	-11.0	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050922\
 Data File : VN072372.D
 Acq On : 09 May 2022 11:41
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: May 10 06:42:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.008	-14.3	90	0.00
50 S	Toluene-d8	1.258	1.374	-9.2	95	0.00
51 T	4-Methyl-2-Pentanone	0.584	0.658	-12.7	95	0.00
52 CM	Toluene	0.864	0.921	-6.6#	92	0.00
53 T	t-1,3-Dichloropropene	0.516	0.609	-18.0	99	0.00
54 T	cis-1,3-Dichloropropene	0.558	0.640	-14.7	97	0.00
55 T	1,1,2-Trichloroethane	0.362	0.391	-8.0	97	0.00
56 T	Ethyl methacrylate	0.537	0.647	-20.5	96	0.00
57 T	1,3-Dichloropropane	0.636	0.683	-7.4	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.129	0.167	-29.5#	94	0.00
59 T	2-Hexanone	0.411	0.473	-15.1	94	0.00
60 T	Dibromochloromethane	0.370	0.426	-15.1	100	0.00
61 T	1,2-Dibromoethane	0.356	0.388	-9.0	97	0.00
62 S	4-Bromofluorobenzene	0.417	0.492	-18.0	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.301	0.327	-8.6	102	0.00
65 PM	Chlorobenzene	1.010	1.065	-5.4	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.381	0.424	-11.3	100	0.00
67 C	Ethyl Benzene	1.762	1.962	-11.4#	95	0.00
68 T	m/p-Xylenes	0.654	0.731	-11.8	95	0.00
69 T	o-Xylene	0.634	0.734	-15.8	96	0.00
70 T	Styrene	0.999	1.198	-19.9	98	0.00
71 P	Bromoform	0.283	0.329	-16.3	102	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	4.416	4.547	-3.0	97	0.00
74 T	N-amyl acetate	1.654	1.784	-7.9	98	0.00
75 P	1,1,2,2-Tetrachloroethane	1.719	1.509	12.2	95	0.00
76 T	1,2,3-Trichloropropane	1.445	1.316	8.9	107	0.00
77 T	Bromobenzene	1.044	1.057	-1.2	103	0.00
78 T	n-propylbenzene	4.865	5.146	-5.8	97	0.00
79 T	2-Chlorotoluene	3.195	3.157	1.2	96	0.00
80 T	1,3,5-Trimethylbenzene	3.678	3.752	-2.0	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.432	0.457	-5.8	96	0.00
82 T	4-Chlorotoluene	2.975	3.028	-1.8	96	0.00
83 T	tert-Butylbenzene	3.229	3.397	-5.2	99	0.00
84 T	1,2,4-Trimethylbenzene	3.558	3.707	-4.2	95	0.00
85 T	sec-Butylbenzene	4.210	4.660	-10.7	101	0.00
86 T	p-Isopropyltoluene	3.282	3.775	-15.0	100	0.00
87 T	1,3-Dichlorobenzene	1.690	1.849	-9.4	104	0.00
88 T	1,4-Dichlorobenzene	1.644	1.810	-10.1	106	0.00
89 T	n-Butylbenzene	2.514	2.991	-19.0	101	0.00
90 T	Hexachloroethane	0.712	0.679	4.6	97	0.00
91 T	1,2-Dichlorobenzene	1.666	1.826	-9.6	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.305	0.308	-1.0	97	0.00
93 T	1,2,4-Trichlorobenzene	0.656	0.860	-31.1#	109	0.00
94 T	Hexachlorobutadiene	0.446	0.468	-4.9	106	0.00
95 T	Naphthalene	1.870	2.709	-44.9#	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050922\
Data File : VN072372.D
Acq On : 09 May 2022 11:41
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: May 10 06:42:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.634	0.860	-35.6#	112	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6