

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050923\
 Data File : VN077634.D
 Acq On : 09 May 2023 16:06
 Operator : JC\MD
 Sample : PB152652TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152652TB

Quant Time: May 10 04:46:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050923W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 09 13:52:31 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	533641	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	963316	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	946633	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	330353	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	329082	48.625	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.240%
35) Dibromofluoromethane	8.172	113	295666	49.212	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.420%
50) Toluene-d8	10.572	98	1196134	52.586	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.180%
62) 4-Bromofluorobenzene	12.854	95	400883	52.108	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.220%
Target Compounds						
					Qvalue	
16) Acetone	4.443	43	41780	25.996	ug/l	96
20) Methylene Chloride	5.284	84	31802	4.775	ug/l #	89
37) Ethyl Acetate	7.566	43	69869	14.738	ug/l	96
43) Isopropyl Acetate	8.695	43	303965	36.569	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050923\
Data File : VN077634.D
Acq On : 09 May 2023 16:06
Operator : JC\MD
Sample : PB152652TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152652TB

Quant Time: May 10 04:46:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050923W.M
Quant Title : SW846 8260
QLast Update : Tue May 09 13:52:31 2023
Response via : Initial Calibration

