

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050923\
 Data File : VN077637.D
 Acq On : 09 May 2023 17:19
 Operator : JC\MD
 Sample : PB152652ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152652ZHE#03

Quant Time: May 10 04:48:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050923W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 09 13:52:31 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	555965	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1046010	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	1007232	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	372791	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	342105	48.519	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.040%
35) Dibromofluoromethane	8.172	113	316734	48.551	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.100%
50) Toluene-d8	10.571	98	1289460	52.207	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.420%
62) 4-Bromofluorobenzene	12.854	95	431197	51.617	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.240%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	41440	24.749	ug/l #	86
20) Methylene Chloride	5.278	84	28493	4.106	ug/l #	83
37) Ethyl Acetate	7.572	43	67155	13.045	ug/l	95
43) Isopropyl Acetate	8.695	43	445190	49.325	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050923\
Data File : VN077637.D
Acq On : 09 May 2023 17:19
Operator : JC\MD
Sample : PB152652ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152652ZHE#03

Quant Time: May 10 04:48:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050923W.M
Quant Title : SW846 8260
QLast Update : Tue May 09 13:52:31 2023
Response via : Initial Calibration

