

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050923\
 Data File : VN077638.D
 Acq On : 09 May 2023 17:43
 Operator : JC\MD
 Sample : PB152652ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152652ZHE#04

Quant Time: May 10 04:48:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050923W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 09 13:52:31 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	549431	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	983083	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	985878	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	353274	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	336795	48.334	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.660%
35) Dibromofluoromethane	8.172	113	311852	50.863	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.720%
50) Toluene-d8	10.572	98	1234076	53.163	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.320%
62) 4-Bromofluorobenzene	12.854	95	414068	52.739	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.480%
Target Compounds						
					Qvalue	
16) Acetone	4.443	43	43165	26.086	ug/l #	87
20) Methylene Chloride	5.284	84	28947	4.221	ug/l	93
37) Ethyl Acetate	7.566	43	64528	13.337	ug/l #	94
43) Isopropyl Acetate	8.695	43	468205	55.195	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050923\
Data File : VN077638.D
Acq On : 09 May 2023 17:43
Operator : JC\MD
Sample : PB152652ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152652ZHE#04

Quant Time: May 10 04:48:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050923W.M
Quant Title : SW846 8260
QLast Update : Tue May 09 13:52:31 2023
Response via : Initial Calibration

