

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050923\
 Data File : VN077643.D
 Acq On : 09 May 2023 19:44
 Operator : JC\MD
 Sample : 02665-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-612-VOC-02

Quant Time: May 10 04:51:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050923W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 09 13:52:31 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	532647	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	964383	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	984634	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	364452	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	325201	48.141	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.280%
35) Dibromofluoromethane	8.172	113	297314	49.432	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.860%
50) Toluene-d8	10.566	98	1252480	55.002	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	110.000%
62) 4-Bromofluorobenzene	12.854	95	425392	55.232	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.460%
Target Compounds						
					Qvalue	
16) Acetone	4.443	43	36002	22.443	ug/l #	85
18) Methyl Acetate	5.043	43	9863	1.551	ug/l #	73
20) Methylene Chloride	5.284	84	37732	5.676	ug/l	91
25) 2-Butanone	7.489	43	29773	12.698	ug/l #	86
37) Ethyl Acetate	7.572	43	61503	12.959	ug/l	98
43) Isopropyl Acetate	8.695	43	279639	33.605	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050923\
Data File : VN077643.D
Acq On : 09 May 2023 19:44
Operator : JC\MD
Sample : 02665-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-612-VOC-02

Quant Time: May 10 04:51:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050923W.M
Quant Title : SW846 8260
QLast Update : Tue May 09 13:52:31 2023
Response via : Initial Calibration

