

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050923\
 Data File : VN077644.D
 Acq On : 09 May 2023 20:08
 Operator : JC\MD
 Sample : 02652-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-DRUMS-OILY-DEBRIS

Quant Time: May 10 04:51:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050923W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 09 13:52:31 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	613983	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1096110	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	1039623	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	407547	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	362658	46.574	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.140%
35) Dibromofluoromethane	8.172	113	335214	49.035	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.080%
50) Toluene-d8	10.572	98	1396704	53.964	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.920%
62) 4-Bromofluorobenzene	12.854	95	444559	50.784	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.560%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.531	59	28050	30.622	ug/l #	86
16) Acetone	4.437	43	77825	42.088	ug/l	100
20) Methylene Chloride	5.278	84	14751	1.925	ug/l #	82
25) 2-Butanone	7.495	43	40908	15.136	ug/l	88
43) Isopropyl Acetate	8.695	43	124577	13.172	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050923\
Data File : VN077644.D
Acq On : 09 May 2023 20:08
Operator : JC\MD
Sample : 02652-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-DRUMS-OILY-DEBRIS

Quant Time: May 10 04:51:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050923W.M
Quant Title : SW846 8260
QLast Update : Tue May 09 13:52:31 2023
Response via : Initial Calibration

