

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050924\
 Data File : VN082048.D
 Acq On : 09 May 2024 14:10
 Operator : JC\MD
 Sample : P2428-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RMW-06B-38-050824

Quant Time: May 10 03:51:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	127132	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	276364	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	247034	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	89712	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	115792	53.966	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.940%	
35) Dibromofluoromethane	8.171	113	84489	50.776	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.560%	
50) Toluene-d8	10.565	98	352536	53.945	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.900%	
62) 4-Bromofluorobenzene	12.847	95	114594	44.359	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 88.720%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050924\
Data File : VN082048.D
Acq On : 09 May 2024 14:10
Operator : JC\MD
Sample : P2428-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RMW-06B-38-050824

Quant Time: May 10 03:51:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

