

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050924\
 Data File : VN082049.D
 Acq On : 09 May 2024 14:34
 Operator : JC\MD
 Sample : PB160759ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB160759ZHE#01

Quant Time: May 10 03:51:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	129891	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.100	114	271526	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	250508	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	92522	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	114173	52.081	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.160%	
35) Dibromofluoromethane	8.171	113	83172	50.875	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.760%	
50) Toluene-d8	10.571	98	349205	54.388	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 108.780%	
62) 4-Bromofluorobenzene	12.847	95	115439	45.482	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 90.960%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	27545	37.851	ug/l	94
20) Methylene Chloride	5.283	84	10044	5.346	ug/l	90
25) 2-Butanone	7.488	43	14744	11.850	ug/l #	87
43) Isopropyl Acetate	8.688	43	251362	46.488	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050924\
Data File : VN082049.D
Acq On : 09 May 2024 14:34
Operator : JC\MD
Sample : PB160759ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB160759ZHE#01

Quant Time: May 10 03:51:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

