

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050924\
 Data File : VN082050.D
 Acq On : 09 May 2024 14:58
 Operator : JC\MD
 Sample : PB160759ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB160759ZHE#02

Quant Time: May 10 03:51:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	124837	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.100	114	271982	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	250992	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	95921	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	113521	53.880	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 107.760%		
35) Dibromofluoromethane	8.171	113	83072	50.729	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.460%		
50) Toluene-d8	10.571	98	357539	55.592	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 111.180%		
62) 4-Bromofluorobenzene	12.847	95	118856	46.750	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 93.500%		
Target Compounds						
16) Acetone	4.430	43	29923	43.512	ug/l	Qvalue 95
20) Methylene Chloride	5.283	84	9869	5.466	ug/l	91
25) 2-Butanone	7.483	43	13710	11.465	ug/l #	66
43) Isopropyl Acetate	8.688	43	274760	50.730	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050924\
Data File : VN082050.D
Acq On : 09 May 2024 14:58
Operator : JC\MD
Sample : PB160759ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB160759ZHE#02

Quant Time: May 10 03:51:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

