

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050924\
 Data File : VN082052.D
 Acq On : 09 May 2024 15:46
 Operator : JC\MD
 Sample : PB160759ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB160759ZHE#04

Quant Time: May 10 03:52:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	122321	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	268767	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	242289	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	89375	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	109690	53.133	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	106.260%	
35) Dibromofluoromethane	8.171	113	79395	49.064	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	98.120%	
50) Toluene-d8	10.571	98	334367	52.611	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	105.220%	
62) 4-Bromofluorobenzene	12.853	95	113494	45.175	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	=	90.340%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	29461	43.748	ug/l	94
20) Methylene Chloride	5.277	84	11028	6.233	ug/l	90
25) 2-Butanone	7.488	43	14386	12.278	ug/l #	71
43) Isopropyl Acetate	8.688	43	273613	51.123	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050924\
Data File : VN082052.D
Acq On : 09 May 2024 15:46
Operator : JC\MD
Sample : PB160759ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB160759ZHE#04

Quant Time: May 10 03:52:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

