

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050924\
 Data File : VN082054.D
 Acq On : 09 May 2024 16:34
 Operator : JC\MD
 Sample : PB160759ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB160759ZHE#06

Quant Time: May 10 03:53:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	127484	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	269387	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	248365	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	88235	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	111728	51.928	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 103.860%		
35) Dibromofluoromethane	8.171	113	83549	51.512	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.020%		
50) Toluene-d8	10.565	98	351888	55.241	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 110.480%		
62) 4-Bromofluorobenzene	12.853	95	114805	45.592	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 91.180%		
Target Compounds						
16) Acetone	4.430	43	25921	36.062	ug/l	98
20) Methylene Chloride	5.277	84	9535	5.171	ug/l #	66
25) 2-Butanone	7.482	43	16997	13.918	ug/l #	81
43) Isopropyl Acetate	8.688	43	260797	48.616	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050924\
Data File : VN082054.D
Acq On : 09 May 2024 16:34
Operator : JC\MD
Sample : PB160759ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB160759ZHE#06

Quant Time: May 10 03:53:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

