

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050924\
 Data File : VN082057.D
 Acq On : 09 May 2024 17:46
 Operator : JC\MD
 Sample : P2413-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-CC

Quant Time: May 10 03:54:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	118771	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.106	114	268830	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	241921	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	92953	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	109880	54.816	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.640%	
35) Dibromofluoromethane	8.165	113	79016	48.818	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.640%	
50) Toluene-d8	10.571	98	342776	53.922	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.840%	
62) 4-Bromofluorobenzene	12.847	95	114463	45.550	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 91.100%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	28519	43.598	ug/l	# 87
20) Methylene Chloride	5.283	84	9042	5.263	ug/l	# 68
25) 2-Butanone	7.483	43	23999	21.094	ug/l	# 81
31) Cyclohexane	8.235	56	8207	2.718	ug/l	# 49
39) Methylcyclohexane	9.600	83	5651	1.767	ug/l	# 85
43) Isopropyl Acetate	8.688	43	222974	41.651	ug/l	# 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050924\
Data File : VN082057.D
Acq On : 09 May 2024 17:46
Operator : JC\MD
Sample : P2413-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-CC

Quant Time: May 10 03:54:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

