

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050924\
 Data File : VN082058.D
 Acq On : 09 May 2024 18:10
 Operator : JC\MD
 Sample : P2413-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-EE

Quant Time: May 10 03:55:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	123411	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.100	114	270214	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	252697	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	93952	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	117401	56.366	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 112.740%		
35) Dibromofluoromethane	8.165	113	85180	52.357	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 104.720%		
50) Toluene-d8	10.571	98	350278	54.820	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 109.640%		
62) 4-Bromofluorobenzene	12.847	95	112607	44.582	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 89.160%		
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	31335	46.423	ug/l	93
20) Methylene Chloride	5.277	84	10942	6.130	ug/l #	69
25) 2-Butanone	7.488	43	23569	19.937	ug/l	94
43) Isopropyl Acetate	8.688	43	257896	47.928	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050924\
Data File : VN082058.D
Acq On : 09 May 2024 18:10
Operator : JC\MD
Sample : P2413-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-EE

Quant Time: May 10 03:55:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

