

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050924\
 Data File : VN082059.D
 Acq On : 09 May 2024 18:34
 Operator : JC\MD
 Sample : P2414-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1011UMR-SS004-0212-01

Quant Time: May 10 03:55:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	129723	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	268502	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	242145	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	90572	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	112422	51.349	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 102.700%	
35) Dibromofluoromethane	8.165	113	79836	49.385	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.760%	
50) Toluene-d8	10.565	98	344121	54.200	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 108.400%	
62) 4-Bromofluorobenzene	12.853	95	111749	44.524	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 89.040%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	27789	38.293	ug/l	94
20) Methylene Chloride	5.277	84	12652	6.743	ug/l #	82
25) 2-Butanone	7.494	43	22238	17.896	ug/l	98
43) Isopropyl Acetate	8.688	43	165637	30.979	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050924\
Data File : VN082059.D
Acq On : 09 May 2024 18:34
Operator : JC\MD
Sample : P2414-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1011UMR-SS004-0212-01

Quant Time: May 10 03:55:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

