

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050924\
 Data File : VN082061.D
 Acq On : 09 May 2024 19:22
 Operator : JC\MD
 Sample : P2398-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-KK

Quant Time: May 10 03:56:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	114922	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	243326	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	219689	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	79494	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	101813	52.492	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 104.980%		
35) Dibromofluoromethane	8.171	113	74544	50.882	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.760%		
50) Toluene-d8	10.565	98	316739	55.049	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 110.100%		
62) 4-Bromofluorobenzene	12.847	95	102175	44.922	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 89.840%		
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	21846	33.351	ug/l	90
20) Methylene Chloride	5.277	84	8563	5.152	ug/l #	73
25) 2-Butanone	7.482	43	19681	17.878	ug/l #	85
43) Isopropyl Acetate	8.688	43	213847	44.133	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050924\
Data File : VN082061.D
Acq On : 09 May 2024 19:22
Operator : JC\MD
Sample : P2398-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-KK

Quant Time: May 10 03:56:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

