

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050925\
 Data File : VN086554.D
 Acq On : 09 May 2025 11:34
 Operator : JC\MD
 Sample : VN0509MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0509MBL01

Quant Time: May 09 23:24:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	161578	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	304753	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	283891	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	123333	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	114071	48.685	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.380%
35) Dibromofluoromethane	8.159	113	89742	63.449	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	126.900%#
50) Toluene-d8	10.565	98	382413	50.589	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.180%
62) 4-Bromofluorobenzene	12.847	95	133637	48.469	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050925\
Data File : VN086554.D
Acq On : 09 May 2025 11:34
Operator : JC\MD
Sample : VN0509MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0509MBL01

Quant Time: May 09 23:24:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

