

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050925\
 Data File : VN086564.D
 Acq On : 09 May 2025 16:08
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: May 09 23:30:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	221112	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	407801	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	394695	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	173381	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	146326	45.637	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.280%
35) Dibromofluoromethane	8.159	113	119857	63.327	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	126.660%#
50) Toluene-d8	10.565	98	515849	50.997	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.000%
62) 4-Bromofluorobenzene	12.847	95	190257	51.568	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050925\
Data File : VN086564.D
Acq On : 09 May 2025 16:08
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: May 09 23:30:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

