

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051019\
 Data File : VN055550.D
 Acq On : 10 May 2019 14:27
 Operator : JC/SP
 Sample : K2758-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IW-20

Quant Time: May 11 01:44:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N050819W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 10 06:48:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	848208	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1433385	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1181783	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	333501	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	576929	45.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.34%	
35) Dibromofluoromethane	7.59	113	451702	48.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.70%	
50) Toluene-d8	10.09	98	1736177	48.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.36%	
62) 4-Bromofluorobenzene	12.40	95	497612	40.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.92%	
Target Compounds						
12) 1,1-Dichloroethene	3.73	96	61819	6.668	ug/l	99
19) Methyl tert-butyl Ether	5.06	73	19186	0.623	ug/l #	84
24) 1,1-Dichloroethane	5.85	63	36727	1.763	ug/l	98
27) cis-1,2-Dichloroethene	6.83	96	118549	10.206	ug/l	91
44) Trichloroethene	8.84	130	96227	9.228	ug/l	96
64) Tetrachloroethene	10.63	164	4153	0.429	ug/l	85
95) Naphthalene	15.14	128	3854	3.198	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN051019\
Data File : VN055550.D
Acq On : 10 May 2019 14:27
Operator : JC/SP
Sample : K2758-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IW-20

Quant Time: May 11 01:44:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N050819W.M
Quant Title : SW846 8260
QLast Update : Fri May 10 06:48:32 2019
Response via : Initial Calibration

