

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051019\
 Data File : VN055558.D
 Acq On : 10 May 2019 17:48
 Operator : JC/SP
 Sample : K2792-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DRUM-COMP

Quant Time: May 11 02:02:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N050819W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 10 06:48:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	840810	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1427494	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1200046	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	382613	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	582016	45.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.94%	
35) Dibromofluoromethane	7.59	113	443634	48.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.36%	
50) Toluene-d8	10.09	98	1764539	49.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.34%	
62) 4-Bromofluorobenzene	12.40	95	520625	43.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.06%	
Target Compounds						
20) Methylene Chloride	4.55	84	178623	14.902	ug/l	99
43) Isopropyl Acetate	8.17	43	28912	1.241	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN051019\
Data File : VN055558.D
Acq On : 10 May 2019 17:48
Operator : JC/SP
Sample : K2792-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DRUM-COMP

Quant Time: May 11 02:02:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N050819W.M
Quant Title : SW846 8260
QLast Update : Fri May 10 06:48:32 2019
Response via : Initial Calibration

