

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051021\
 Data File : VN067071.D
 Acq On : 10 May 2021 15:01
 Operator : JC/MD
 Sample : M2323-04 20X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AST2-COPR3-1-6

Manual Integrations
 APPROVED

MMDadoda
 5/11/2021 10:37:16 AM

Quant Time: May 11 01:47:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.581	168	326130	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	8.509	114	537756	50.00	ug/l	-0.01
63) Chlorobenzene-d5	11.350	117	459240	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.294	152	153098	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.943	65	287302	55.97	ug/l	-0.01
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.94%
35) Dibromofluoromethane	7.504	113	197875	54.19	ug/l	-0.01
Spiked Amount	50.000	Range	69 - 133	Recovery	=	108.38%
50) Toluene-d8	10.025	98	727106	52.29	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.58%
62) 4-Bromofluorobenzene	12.350	95	220573m	46.84	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.68%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051021\
Data File : VN067071.D
Acq On : 10 May 2021 15:01
Operator : JC/MD
Sample : M2323-04 20X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AST2-COPR3-1-6

Manual Integrations
APPROVED

MMDadoda
5/11/2021 10:37:16 AM

Quant Time: May 11 01:47:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

