

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051021\
 Data File : VN067072.D
 Acq On : 10 May 2021 15:25
 Operator : JC/MD
 Sample : PB136216ZHE#03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB136216ZHE#03

Quant Time: May 11 01:47:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.584	168	343228	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.512	114	574361	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	529660	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.291	152	149141	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.946	65	297825	55.13	ug/l	0.00
Spiked Amount 50.000	Range 61	- 141	Recovery	=	110.26%	
35) Dibromofluoromethane	7.504	113	212265	54.42	ug/l	-0.01
Spiked Amount 50.000	Range 69	- 133	Recovery	=	108.84%	
50) Toluene-d8	10.027	98	797827	53.71	ug/l	0.00
Spiked Amount 50.000	Range 65	- 126	Recovery	=	107.42%	
62) 4-Bromofluorobenzene	12.347	95	226064	44.95	ug/l	0.00
Spiked Amount 50.000	Range 58	- 135	Recovery	=	89.90%	
Target Compounds						Qvalue
16) Acetone	3.741	43	76013	40.00	ug/l	100
20) Methylene Chloride	4.460	84	123856	25.69	ug/l	92
43) Isopropyl Acetate	8.088	43	37957	3.83	ug/l	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051021\
Data File : VN067072.D
Acq On : 10 May 2021 15:25
Operator : JC/MD
Sample : PB136216ZHE#03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB136216ZHE#03

Quant Time: May 11 01:47:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

