

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051021\
 Data File : VN067073.D
 Acq On : 10 May 2021 15:47
 Operator : JC/MD
 Sample : PB136216ZHE#04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB136216ZHE#04

Quant Time: May 11 01:47:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.584	168	345397	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.512	114	580197	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	543087	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.292	152	183457	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.946	65	302299	55.60	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.20%
35) Dibromofluoromethane	7.507	113	213284	54.13	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	108.26%
50) Toluene-d8	10.025	98	808970	53.92	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.84%
62) 4-Bromofluorobenzene	12.347	95	236542	46.56	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.12%
Target Compounds						
16) Acetone	3.738	43	69290	36.23	ug/l	95
20) Methylene Chloride	4.457	84	116197	23.95	ug/l	93
43) Isopropyl Acetate	8.088	43	38079	3.81	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051021\
Data File : VN067073.D
Acq On : 10 May 2021 15:47
Operator : JC/MD
Sample : PB136216ZHE#04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB136216ZHE#04

Quant Time: May 11 01:47:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

