

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051021\
 Data File : VN067075.D
 Acq On : 10 May 2021 16:32
 Operator : JC/MD
 Sample : M2323-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AST2-COPR3-1-6

Manual Integrations
 APPROVED

MMDadoda
 5/11/2021 10:37:18 AM

Quant Time: May 11 01:48:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.587	168	340607	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.515	114	570837	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	505576	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.294	152	166401	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	295340	55.09	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 110.18%			
35) Dibromofluoromethane	7.509	113	213127	54.98	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 109.96%			
50) Toluene-d8	10.027	98	787098	53.32	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 106.64%			
62) 4-Bromofluorobenzene	12.350	95	221736	44.36	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 88.72%			
Target Compounds						
16) Acetone	3.746	43	38386	20.35	ug/l	Qvalue 95
18) Methyl Acetate	4.240	43	8741m	1.62	ug/l	
20) Methylene Chloride	4.462	84	34928	7.30	ug/l	96
43) Isopropyl Acetate	8.091	43	20248	2.06	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051021\
Data File : VN067075.D
Acq On : 10 May 2021 16:32
Operator : JC/MD
Sample : M2323-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AST2-COPR3-1-6

Manual Integrations
APPROVED

MMDadoda
5/11/2021 10:37:18 AM

Quant Time: May 11 01:48:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

