

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051021\
 Data File : VN067077.D
 Acq On : 10 May 2021 17:19
 Operator : JC/MD
 Sample : M2295-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DRUM-A-B-C

Manual Integrations
 APPROVED

MMDadoda
 5/11/2021 10:37:23 AM

Quant Time: May 11 01:48:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.584	168	339389	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.515	114	571688	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	535106	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.294	152	183420	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.946	65	298731	55.92	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 111.84%	
35) Dibromofluoromethane	7.506	113	212058	54.62	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 109.24%	
50) Toluene-d8	10.027	98	790519	53.47	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 106.94%	
62) 4-Bromofluorobenzene	12.347	95	229883	45.92	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 91.84%	
Target Compounds						
16) Acetone	3.741	43	66887	35.59	ug/l	Qvalue 94
18) Methyl Acetate	4.250	43	6890m	1.28	ug/l	
20) Methylene Chloride	4.462	84	60595	12.71	ug/l	95
43) Isopropyl Acetate	8.088	43	34452	3.50	ug/l	# 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051021\
Data File : VN067077.D
Acq On : 10 May 2021 17:19
Operator : JC/MD
Sample : M2295-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DRUM-A-B-C

Manual Integrations
APPROVED

MMDadoda
5/11/2021 10:37:23 AM

Quant Time: May 11 01:48:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

