

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051021\
 Data File : VN067078.D
 Acq On : 10 May 2021 17:42
 Operator : JC/MD
 Sample : M2295-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DRUM-D-E

Quant Time: May 11 01:48:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.584	168	347486	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	579240	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	551085	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.291	152	192760	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.946	65	298949	54.66	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	109.32%
35) Dibromofluoromethane	7.504	113	204332	51.95	ug/l	-0.01
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.90%
50) Toluene-d8	10.027	98	830888	55.47	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	110.94%
62) 4-Bromofluorobenzene	12.347	95	244956	48.30	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.60%
Target Compounds						
16) Acetone	3.741	43	63105	32.80	ug/l	99
20) Methylene Chloride	4.462	84	66134	13.55	ug/l	95
43) Isopropyl Acetate	8.086	43	27278	2.73	ug/l #	88
95) Naphthalene	15.091	128	42981	5.08	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051021\
Data File : VN067078.D
Acq On : 10 May 2021 17:42
Operator : JC/MD
Sample : M2295-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DRUM-D-E

Quant Time: May 11 01:48:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

